

Makrite Industries Inc.

N95 & N99 Series Particulate Respirator



Data Sheet

Main Features

The N95 & N99 Series are extremely durable with a soft and comfortable inner surface, with an adjustable nose piece and secured headstraps to provide a proper fit.

Designations

- N = Non-oil
- R = Oil Resistance
- P = Proof Oil


Standards

This product has been tested and meets the NIOSH standard of N95 & N99. Approved by NIOSH. The main performance tests in this standard are:

- * Filter Penetration
- * Breathing Resistance

Storage and Transportation

The N95 Series Particulate Respirators have a shelf life of 3 years. End of shelf life is marked on the product packaging. Before initial use, always check that the product is within the stated shelf life (use by date). Product should be stored in clean, dry conditions within the temperature range: -5°C to + 50°C with a maximum relative humidity of <75%. When storing or transporting this product use original packaging provided.



Materials

	9500-N95 9500-N95S 9500P-N95	9500V-N95 9500AV-N95	MK910-N95 MK910-N95S	MKN95-910V MKN95-910PV	910-N95VOV	N99-810V	9500-N95OV
Welding Straps	●	●	●	●			●
Staple Straps					●	●	
Nose Foam	●	●	●	●	●	●	●
Nose Clip	●	●	●	●	●	●	●
Static Charged Filter	●	●	●	●	●	●	●
Exhalation Valve		●		●	●	●	
Carbon					●		●

Makrite Industries Inc.

Certificate Number

ITEM NO.	Certificate No.
9500-N95	TC-84A-5411
9500-N95S	TC-84A-5463
9500P-N95	TC-84A-5466
9500V-N95	TC-84A-5460
9500AV-N95	TC-84A-5639
MK910-N95	TC-84A-5408
MK910-N95S	TC-84A-5464
MKN95-910V	TC-84A-5409
MKN95-910PV	TC-84A-5472
910-N95VOV	TC-84A-4644
N99-810V	TC-84A-4081
9500-N95OV	TC-84A-5525

LIMITED WARRANTY IMPORTANT NOTICE TO PURCHASER :

This limited warranty is made in lieu of the warranties of merchantability, fitness for particular purposes and all other warranties, express or implied. There are no other warranties which extend beyond the description on the face herof. The physical standards and specifications of Makrite will be met by products sold.

Exclusive Remedies: damages for the breach of this limited warranty are limited to the replacement of such quantity of Makrite products proved to be defectively manufactured. Except as provided above, Makrite shall not be liable or responsible for any loss, damage, or liability, direct, indirect, incidental, special or consequential, arising out of sale, use, misuse or the inability to use products by the user.

Makrite Industries Inc.

WARNINGS :

1. Wearer must read fitting instructions and use limitations before use.
2. This respirator does not protect against the risk of contracting disease or infection.
3. Individuals highly sensitive to natural rubber latex may have an allergic reaction.
4. Wearer must done fit test either quantitatively or qualitatively by a qualified fit tester before using the respirator in a contaminated environment, as per OSHA guidelines, 29 CFR 1910.134.

SERVICE TIME LIMITATION

1. Do not use the same respirator for more than 8 hours, continuous or intermittent, in dirty workplaces that could result in high filter loading.
2. Leave contaminated area and remove respirator if breathing becomes difficult or proper fit cannot be obtained or the respirator is damaged, soiled, or distorted. Do not use the same respirator again once removed.
3. The maximum service time of this respirator is five (5) consecutive calendar days (including days of non-use), beginning from the first day of use.

Respirator Fitting Instructions



1 . Hold the respirator in hand with the nosepiece at your fingertips, allowing the headbands to hang freely below your hand.

2 . Press the respirator firmly against your face with the nosepiece on the bridge of your nose.

3 . Stretch and position the top band high on the back of the head. Stretch the bottom band over the head and position below your ears.

4 . Using both hands, mold nosepiece to the shape of your nose.

5 . To test fit: a) Cup both hands over the respirator being careful not to disturb position, and b) inhale vigorously. If air leaks around the edges, reposition the straps or adjust strap tension for better fit. Please carefully follow these fitting instructions during each use to achieve proper fit.