

Distributed by:



800-826-8302 nelsonjameson.com

## CERTIFICATE OF ANALYSIS

This is to verify that all items noted below are in conformance with the Purchase Order Contract, drawings, specifications, and other applicable documentation; that all process certifications, material certifications, are on file at this or another Bel-Art facility and are available for review by the customer and or Government Representatives upon request.

**Product #:** 28994-0000 **Mfg. Date:** N/A **Lot #:** GENERAL

**Description:** PITCHER,PP,GRADUATED,5000ML.

**Material:** Polypropylene

**Dimensions:** Top OD including spout and handle 10.125", w/o spout and handle 8.25", Bottom OD 6.5", Height 10.75"  
ID at top 7", bottom 6.25", Ht 9.875", Spout .75"  
Wall Thickness .079

**Other:** Graduations on one side in 100ml increments, Accuracy = +/- 2%  
not affected by autoclaving , Steam Autoclavable @ 121°C  
Resin meets FDA 21 CFR 177.1520.  
To the best of our knowledge based on information provided to us,  
The materials used do not contain nor have come in contact with  
any animal origins. Based on the process and materials involved in the  
manufacturing of this product, this product is considered to be BSE/TSE  
protein free. Based on the information provided to us this product does  
not contain or come in contact with lead, latex, or other contaminants.

Manufactured by: SP Scienceware - BelArt  
2024 Broad Street Pocomoke City, MD 21851  
An ISO-9001:2015 facility (certificate US19/819943409)  
Made in the USA

**Signature:** 

**Printed Name:** Nancy K. Trogani

**Title:** Technical Support Analyst

**Date:** 5/11/2021

Distributed by:



800-826-8302 nelsonjameson.com

FTS | Genevac | Hotpack | Hull | i-Dositecno | Service | VirTis | Wilmad-LabGlass