

WaterWorks™

Free Chlorine Check

Ultra High



Part Number 480024

Contains 50 Test Strips

Tests Free Chlorine from 0–750 ppm

Interferences: Similar color development may occur if other oxidizing agents, such as Chlorine Dioxide, are present in the sample.



For Technical Assistance
Call: 1-803-329-0162
SENSAFE.COM




IMPORTANT: Store in cool, dry place (below 90°F/32°C) and away from direct sunlight.

TEST PROCEDURE:

1. Dip one test strip into a water sample for **1 second**.
2. Remove the strip and shake once*, briskly, to remove excess water.
3. **Wait 30 seconds** (for values above 400ppm, wait 90 seconds for full color development), then match to the color chart. Complete color matching **within 30 seconds**.

*Shake strip towards ground and away from possible eye contact.



US PATENT 5491094
WATERWORKS™ is a trademark of
Industrial Test Systems, Inc., Rock Hill, SC USA
©2015 Industrial Test Systems, Inc. R1015AGA-FCUH