

3M™ Petrifilm™ Yeast and Mold Count Plates

Product Information

MEDIUM:	3M Petrifilm Yeast and Mold Count Plates (6407/6417/6445)																												
ISO/BSI CERTIFICATE OF REGISTRATION NUMBER:	Manufactured at Brookings, SD, USA (ISO 9001:2008, FM 14552)																												
DATE OF EXPIRATION/ LOT NUMBER:	Expiry and lot number indicated on each package. Lot number indicated on each plate.																												
FORMULATION:	Nutrients supplemented with antibiotics, a cold-water soluble gel, and indicator																												
METHOD OF PREPARATION:	Nutrients and gels coated onto film. For use, hydrate with one mL aqueous sample or dilution of sample. See product instructions for detailed instructions.																												
CONTAMINATION CHECK:	96 randomized plates per batch tested Incubated at 25°C for 5 days Utilizes statistical sampling plan																												
EFFICACY CHECK:	<table><tr><th colspan="2">Organisms tested include:</th><th>Acceptable Batch</th></tr><tr><td><u>Organism</u></td><td></td><td></td></tr><tr><td><i>Enterococcus faecalis</i></td><td></td><td>No Growth</td></tr><tr><td><i>Acinetobacter</i> species</td><td></td><td>No Growth</td></tr><tr><td><i>Aspergillus niger</i></td><td></td><td>Productivity Ratio ≥ 0.7</td></tr><tr><td><i>Paecilomyces</i> species</td><td></td><td>Productivity Ratio ≥ 0.7</td></tr><tr><td><i>Botrytis</i> species</td><td></td><td>Productivity Ratio ≥ 0.7</td></tr><tr><td><i>Hansenula anomala</i></td><td></td><td>Productivity Ratio ≥ 0.7</td></tr><tr><td><i>Saccharomyces cerevisiae</i></td><td></td><td>Productivity Ratio ≥ 0.7</td></tr></table>		Organisms tested include:		Acceptable Batch	<u>Organism</u>			<i>Enterococcus faecalis</i>		No Growth	<i>Acinetobacter</i> species		No Growth	<i>Aspergillus niger</i>		Productivity Ratio ≥ 0.7	<i>Paecilomyces</i> species		Productivity Ratio ≥ 0.7	<i>Botrytis</i> species		Productivity Ratio ≥ 0.7	<i>Hansenula anomala</i>		Productivity Ratio ≥ 0.7	<i>Saccharomyces cerevisiae</i>		Productivity Ratio ≥ 0.7
Organisms tested include:		Acceptable Batch																											
<u>Organism</u>																													
<i>Enterococcus faecalis</i>		No Growth																											
<i>Acinetobacter</i> species		No Growth																											
<i>Aspergillus niger</i>		Productivity Ratio ≥ 0.7																											
<i>Paecilomyces</i> species		Productivity Ratio ≥ 0.7																											
<i>Botrytis</i> species		Productivity Ratio ≥ 0.7																											
<i>Hansenula anomala</i>		Productivity Ratio ≥ 0.7																											
<i>Saccharomyces cerevisiae</i>		Productivity Ratio ≥ 0.7																											
PACKAGING:	Pack size: 50 plates per foil pouch Film grade: Plastic foil laminate Seal integrity check: Pressure decay test performed																												
MEASUREMENT SYSTEMS CALIBRATION AND TRACEABILITY:	Incubator temperature 3M internal calibration Minimum calibration once per year for all equipment																												
MEDIA QUALITY STATEMENT:	Certificate of Analysis available on 3m.com/foodsafety																												
SHELF LIFE:	18 months from date of manufacture																												
STORAGE CONDITIONS:	Store at temperatures less than or equal to 8°C																												

Distributed by:

NELSON JAMESON
INC.

800-826-8302 nelsonjameson.com