



8

WaterWorks™

Free Chlorine Check

Ultra High II

**Part Number 480124****Contains Test Strips**

Tests Free Chlorine from 0–2000 ppm

KEEP OUT OF REACH OF CHILDREN

Colors calibrated at 75–80°F (24–27°C).

Interferences: Similar color development may occur if other oxidizing agents, such as Chlorine Dioxide, are present in the sample.



For Technical Assistance
Call: 1-803-329-0162
SENSAFE.COM



IMPORTANT: Store in cool, dry place (below 90°F/32°C) and away from direct sunlight.

TEST PROCEDURE:

1. Dip one test strip into a water sample for **1 second**.
2. Remove the strip and shake once*, briskly, to remove excess water.
3. **Wait 1 Minute**, then match to the color chart below. Complete color matching **within 15 seconds**.

Note: Color on the pad may not be uniform, this is normal.

*Shake strip towards ground and away from possible eye contact.



US Patent # 5491094

WATERWORKS™ is a trademark of

Industrial Test Systems, Inc., Rock Hill, SC USA

©2015 Industrial Test Systems, Inc. R1015AGA-FCUH2