




Peracetic
Acid
ppm



TO USE:

Using a cup-size sample, immerse test strip and remove with **pad face up. DO NOT SHAKE OFF EXCESS WATER.** Wait 10 seconds and compare to color chart immediately.



TIPS

- Keep wet fingers out of the vial.
- Close vial tightly after removing strip.
- Store in a cool, dry place.

 **LaMotte**
Chestertown, MD 21620
www.lamotte.com

made in
USA 

Code 3000HR

