



Delvotest®

Delvotest® ampoules photo instructions



1. Cut off the required number of ampoules with a pair of scissors. Be careful not to damage the foil of remaining ampoules.




2. Open ampoules by punching a hole in the aluminium foil with the syringe. Mark the ampoules for sample identification.



3. Add one nutrient tablet to each ampoule upon the solid agar with the tweezers. Do not touch the tablets with your fingers.



4. Attach a new disposable pipette to the syringe. Do not touch the tip-end, which will be in contact with the milk.




5. Depress the plunger completely, dip the tip in the milk sample and allow the plunger to return slowly under pressure of the spring.



6. Empty the syringe into the corresponding marked ampoule by slowly depressing the plunger of the syringe. Use a fresh disposable pipette for each milk sample.



7. Check the temperature of the incubator ($64^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$). Put the ampoules into the incubator. Record the time and set timer for: **P** ampoules 2,5 hours or for **SP** ampoules 3 hours or use control time.




8. Read the colour of the lower 2/3 part of the solid agar in the ampoules after the required incubation time.

Delvotest® reading colours

Negative



Detection limit



Positive



Distributed by:

NELSON JAMESON INC.

800-826-8302 nelsonjameson.com

Although diligent care has been used to ensure that the information provided herein is accurate, nothing contained herein can be construed to imply any representation or warranty as to the accuracy, currency or completeness of this information. The content of this document is subject to change without further notice. This document is non-controlled and will not be automatically replaced when changed. Please contact us for the latest version of this document or further information.
DSM Food Specialties B.V. | A. Fleminglaan 1 | 2613 AX Delft | The Netherlands